



University of Colorado Hospital

ANSCHUTZ MEDICAL CAMPUS

Test Update Molecular Diagnostic Sensitivity Alert

The UCH Molecular Diagnostics Laboratory has recently become aware of decreases in quantitative testing sensitivities due to a specific reagent purchased from a vendor. We are partnering with the vendor in hopes of isolating the problem. In the meantime, please note the temporary changes to testing sensitivities indicated below.

| Test Description | Previous Sensitivity | New Sensitivity |
|-------------------------|-----------------------------|------------------------|
| Adenovirus PCR | 100 copies/mL | 200 copies/mL |
| BK Virus PCR | 600 copies/mL | 1000 copies/mL |
| CMV PCR | 100 copies/mL | 250 copies/mL |
| EBV PCR | 100 copies/mL | 500 copies/mL |
| HHV6 PCR | 250 copies/mL | 500 copies/mL |
| JC Virus PCR | 100 copies/mL | 250 copies/mL |

If you have any questions or special concerns, contact the University of Colorado Hospital Clinical Laboratory Molecular Diagnostics Department at 720-848-6892.

Issued 12/20/2011

Clinical Laboratory
Medical Director Ronald B. Lepoff, MD